REV 1
CHANGE
3/2/98

RADIATION CO[illegible] GROUP
RADIOLOGIC/ [illegible]RVEY FORM



| Area (map reference if applicable) | Accelerator Operating Conditions | Instrument(s) FAG |
|---|---|---|
| Hall-C | 75μA 1 PASS CW | Serial # 10585 Cal due 7-5-01 |

Reason for survey: CONTROLLED ACCESS

1. Post all areas that are >5 mR/hr as a radiation area
2. Annotate the precise location of radiation areas
3. Annotate any abnormal radiation levels

Legend

All readings in mr/hr whole body (unless annotated)

——— Denotes posted area

(#) Denotes smear location

(refer to page 2 for smear results)

Contact dose rate
WB dose rate
Item description




| Summary | | |
|---|---|---|
| Types of Areas (Circle all that apply) | | |
| RCA | Rad | High Rad Cont |
| All areas posted | Y | N |
| Highest WB Dose Rate (mr/hr) | | |
| RCA | Rad | High Rad 1.2 R/hr |

| For Beam Enclosure Surveys | | | |
|---|---|---|---|
| Initial Entry Survey | Y | N | |
| Access mode | Swp | Cont | Rest |
| Continuous Rad Escort | Y | N | |

| Performed by: | Date/Time: | Crew Chief Review: | RCG Review: |
|---|---|---|---|
| [illegible] | 3-23-01 0915 | [illegible] | [illegible] |

CHANGE: N/A
3/2/98

RADIATION CONTROL GROUP
RADIOLOGICAL SURVEY FORM

PAGE 1 OF 1

| Area (map reference if applicable) | Accelerator Operating Conditions | Instrument(s) 611B Telefector |
|---|---|---|
| Hall-C | 100 nA 2pass CW | Serial # 101891 Cal due 10-16-01 |

Reason for survey: Work in G-0 Tent

1. Post all areas that are >5 mR/hr as a radiation area
2. Annotate the precise location of radiation areas
3. Annotate any abnormal radiation levels

Legend

All readings in mr/hr whole body (unless annotated)

-------- Denotes posted area

(#) Denotes smear location (refer to page 2 for smear results)

Contact dose rate
WB dose rate
Item description




| Summary | | |
|---|---|---|
| Types of Areas (Circle all that apply) | | |
| RCA Rad High Rad Cont | | |
| All areas posted Y N | | |
| Highest WB Dose Rate (mr/hr) | | |
| RCA | Rad | High Rad |
| 4 | 35 | / |

| For Beam Enclosure Surveys |
|---|
| Initial Entry Survey Y N |
| Access mode Swp Cont Rest |
| Continuous Rad Escort Y N |

* At target Chamber level

| Performed by: | Date/Time: | Crew Chief Review: | RCG Review: |
|---|---|---|---|
| [illegible] | 9-10-01 / 0930 | [illegible] | [illegible] |